

## ReadMe\_MultiAnalytesMU.pdf:-

### Outline documentation for MS Excel and Apple Notes Spreadsheets

---

#### Purpose

The revised ISO/IEC 17025:2017 standard requires Measurement Uncertainty (MU) of all chemical test results to be available. The Australian National Association of Testing Authorities (NATA) requirements for this standard are described in the NATA General Accreditation Guidance documents <sup>{1}</sup>.

The spreadsheet may be useful to assist Laboratory Analysts and Assessors to ensure compliance.

---

#### Background

Two approaches to establishing a new test method are to use MU "bottom up" evaluations, which are often tedious and difficult to evaluate, or a "top down" approach for a quicker determination of repeatability, reproducibility and trueness of a test, where: "accuracy = trueness (estimates of bias) + precision (estimates of random variability)".

Volumetric/mass uncertainties can be much smaller than those from sample digestion/extraction/clean-up, or those from even instrument variability. In these cases the bottom up approach is often not suitable. The guidance "simpler is often better" may explain why techniques such as QuEChERS (Quick, Easy, Cheap, Effective, Rugged and Safe) can produce excellent results with a wide range of organic analytes; or ICP-MS with quite "dirty" inorganic sample digests may be better than complex multi-step clean ups.

Suitable reference materials and proficiency tests are not always available to establish methods for new analyte/matrix combinations. An established "top down" technique is to spike "blank" matrices with analytes, and use recovery and statistical information to establish fitness of purpose. A useful indicator is the Horwitz Ratio (HorRat) <sup>{2}</sup> - The ratio of observed relative standard deviation calculated from actual performance data, to a prediction calculated from the Horwitz equation at a particular concentration. The ratio can be used to define "Recommended", "Acceptable (but requires an explanation)" and "Not Acceptable" ratios <sup>{4}</sup>. Techniques such as mass spectrometry with isotopic spiking can produce better "low level" results than expected, these are displayed for checking. Checking for anomalous test "Outliers" may improve Method Detection Limits. This spreadsheet uses Grubbs' Test <sup>{3}</sup> for each set.

Sometimes hundreds of results must be evaluated for each test - Automation of Horwitz ratio tests may be a good initial approach. The spreadsheet can be used for up to 24 analytes - Each with up to 32 results. The author has used these techniques in spreadsheets over a number of years to evaluate many methods. After development and combination of these basic tools I hope that they may be useful to others.

---

#### General

The system uses only "cell formula" and has no "macros", "program code", or external links (e.g. to webpages or email). **It should be "safe" to distribute.** The sheets are [\[MainTable\]](#), [\[SampleResults\]](#), [\[Distributions\]](#) and [\[Units\]](#), the last two are "look-ups". The last sheet [\[Notes\]](#) is a quick users' guide.

There are two versions:-

**1:** Apple "Numbers" - Can be used with an iPad; a large iPhone (Or a smaller one, connected via the Apple HDMI adapter to an external monitor/TV or the Apple TV and/or a Bluetooth keyboard); an Apple Mac Computer; or, via an iCloud account on any Windows PC with a suitable internet browser. Current versions of the Apple spreadsheet program do not support "cell-level" locking - Some care is needed to avoid deleting data if a field is accidentally selected, and so "locked shapes" are positioned over them.

**2:** Microsoft Excel - Can be run on any modern PC, or in the Cloud with Office 365. Excel supports cell-level locking, and the spreadsheet's relevant cells are protected. The password to unlock the whole spreadsheet is "**timstrutt**" (no "quotes").

For speed and simplicity, much of the work in creating the original spreadsheet was done using Apple Notes version (Versions 6 to 13 - Often on an iPad). These were then ported to the Excel version and required a small amount of further debugging. The main exception was that cell "Conditional Formatting" colours and backgrounds often varied considerably.

[Descriptions and the Use of Individual Sheets Follows Below...](#)

## MultiAnalytesMU Spreadsheet - [Notes]

This is the last sheet. It gives an overview of the spreadsheets, and can be used as a reference.

Apple Numbers Spreadsheet Tab

+	MainTable	SampleResults	Distributions	Units	Notes
<p>This spreadsheet may be useful to assist Laboratory Analysts and Assessors ensure compliance with ISO/ IEC 17025:2017. The revised standard requires Measurement Uncertainty (MU) of chemical test results to be available. It is designed to help produce necessary information by using sample blanks, spikes, and comparisons with reference materials. Sample "Low Spikes" could be used to determine Recoveries and Method Detection Limits. Spikes establish MUs at different concentrations with different matrices. It may be helpful when developing new methods or for "novel" analytes and matrices.</p> <p>Created with Apple "Numbers" Version 13, and Microsoft "Excel" 2013, spreadsheet programs - Using only "Cell Functions". No "Code", Macros, or Hyperlinks are used - <b>It should be "Safe" to distribute...</b></p> <p>The spreadsheet uses some <b>conventions</b>: In [MainTable] cells where the user is expected to input data have <b>blue text</b> : labels, <b>bold blue</b> indicates that a selection is made from a list. Default look-ups are <b>grey</b>. Checks have [<b>pale backgrounds</b>]; Warnings have [<b>darker backgrounds</b>]; and failure/out-of-range warnings have [<b>red backgrounds</b>]; Replaced MDLs or Significances have [<b>blue backgrounds</b>] <b>Important data is purple</b>.</p> <p><b>Getting Started:</b> The sheets are [MainTable], [SampleResults], [Distributions], [Units], and [<b>Notes</b>]. The last three are "look-ups" or information. <b>Data should be added or changed as outlined below...</b></p> <p><b>[MainTable]:</b> Used to set the parameters of the Test:-            The Test Name that was used to create the results is a "look-up" from the [SamplesResults] sheet.            The Analyte Name is a "look-up" from the [SamplesResults] sheet.            Select <b>Report Type</b> from the drop-down - This will produce different a different [Report] for each type.            Select <b>Units</b> from the drop-down - This will change the Horwitz Ratio (larger concentrations should give higher precision).            Enter '<b>Spike Value</b>' for each column (Values between 0.00001 and 10,000,000).            MDL (From Mean of Previous Low Spikes) - If a value is entered, it will be used instead of the Results calculation.            Select <b>Number of Significant figures</b> - Usually <b>2 - 1</b> may be useful if near Quantitation limits, 3 for special precision.            Horwitz and Outlier fields with coloured backgrounds are not editable, but require consideration - Outliers or errors may well be the cause...  <b>Now go to the [SampleResults] Sheet to check/update outliers as below...</b></p> <p><b>[SampleResults]:</b> Used to add SampleID and Results. There are a number of columns:-            The First Column is used to Add/Edit <b>SampleID</b> (or Paste from another application). <b>The 1st Top (Header) row must contain the Test Name.</b>  <b>Other Header rows must contain the Analyte Name.</b> The other columns are for <b>plain numerical Results</b> (or Paste from another application).            Possible result <b>Outliers</b> can be indicated by adding text e.g. Changing 77 to <b>?77</b> or <b>77!</b> Adding text to Outliers will cause it to be ignored by the Horwitz and Grubbs calculations. If pasted results are not plain numbers, cell formatting will be lost.            The sheet can have many rows, but normally 3-11 are used (5, or more recommended).</p> <p><b>[Distributions]:</b> Used to calculate Method Detection Limit (t-test) and Grubbs' critical value for 0.05 (95%) level for 1-tail.            This is not editable. It is recommended that the user consults the NIST and RSC papers, referenced in the text at the bottom of the sheet, for further background information.</p> <p><b>[Units]:</b> Used to calculate scale of Horwitz values, and in [<b>Report</b>] and [<b>Calculations - Main Table</b>]-            It is normally Locked, but any changes must be reflected in [<b>MU Calculations</b>] "<b>Units</b>" drop-down.</p> <p><b>For further information, please consult these References:-</b>            Planning and Reporting Method Validation Studies; Supplement to EURACHEM Guide on the Fitness for Purpose of Analytical Methods, First edition 2019            Blanks in Method Validation, Supplement to EURACHEM Guide The Fitness for Purpose of Analytical Methods, First Edition 2019</p> <p>The <b>password</b> to unlock individual MS Excel sheets and cells is "timstrutt" (no "quotes"). Apple Sheets use Locked "Shapes".</p> <p>The [<b>MainTable</b>] uses "hidden" rows with a <b>red font</b> - <b>Changing or deleting them will cause breakages...</b></p> <p><b>Warning:</b> The calculations are based upon standard literature methods, but some approximations have been used. See the references for background information.</p> <p>This spread-sheet is "Public Domain". Anyone is free to copy, modify, publish, use, or distribute it for any purpose, commercial or non-commercial, and by any means. The author is retired; carries out "Pro Bono" work only; and does not accept any responsibility for any loss caused by the use of, or failure to use, this spreadsheet. Please do not remove my name, copyright &amp; version information.</p> <p style="text-align: right;"><b>Version: 2023-06-010 - ©Tim Strutt.</b></p>					

←	→	MainTable	SampleResults	Distributions	Units	Notes	+	:	←
<p>READY    </p>									

Microsoft Excel Spreadsheet Tab

### MultiAnalytesMU Spreadsheet - [MainTable]

This is the first sheet. It shows calculations for the selected Report Type. Normally 24 Analyte columns are shown for up to 32 Results in the [SampleResults] sheet. Ideally, a white background for the Horwitz Ratio field will be seen. Red backgrounds are indications that the spike level is much higher/lower than the result data, or the result is much higher/lower than expected.

Test Name (TopRow, 1st Column in SampleResults)	ICP-MS	ICP-MS	ICP-MS	ICP-MS	ICP-MS
Analyte (Top Row in SampleResults Sheet)	Al	As	B	Ba	Be
Select Report Type (Spike, Ref...):	Low Spike	Low Spike	Low Spike	Low Spike	Low Spike
Select Units (ppm, µg/kg, g/L...):	µg/kg	µg/kg	µg/kg	µg/kg	µg/kg
Select # of Significant Figures:	2	2	2	2	2
Enter Expected Analyte Concentration For Analytes:	90	2.2	99	90	92
MDL - Optional (From Mean of Previous Low Spikes):					
Total Number of Results for Analyte	9 (1 Excluded!)	9	9	9	9
RSD% (CV)	7.7%	8.8%	4.2%	16.3%	143.7%
bias as %age (100 - Recovery)	7.6%	27.0%	5.7%	-12.4%	-73.5%
Horwitz Ratio (Ideally 0.35 ≤ HorRat ≤ 1) Check: Very High	0.35	0.40	0.19	0.74	6.53
Grubbs' Outlier Check: 3 Found?				140.1	771
Result to 2 Significant figures	83	1.6	93	100	160
MDL = (MDL Factor x SD)	21	0.6	12	42	383
PQL (Practical Quantitation limit)	62	1.7	36	130	1100
% Recovery	92%	73%	94%	112%	173%
Corrected for Recovery with Uncertainty	90 ± 21 µg/kg	2.2 ± 0.6 µg/kg	99 ± 12 µg/kg	90 ± 42 µg/kg	92 ± 383 µg/kg

This shows at least one analyte has a very high HorRat (red background) and has found 3 possible Grubbs outliers.

This has an excluded result. The highlighted result below has been edited by adding "!" Which makes it "non-numeric".

Check if the "Low" HorRat is "Acceptable" - OK if test is highly developed/automated.

"Very High" HorRat - Typographical error, Result entered as "771" instead of correct "77.1"

Calculations can be archived by printing, creating a PDF using the software's "print" facility and saving the resulting file in a suitable folder, or archiving the entire (renamed?) spreadsheet in its own folder.

### MultiAnalytesMU Spreadsheet - [SampleResults]

This is the second sheet. This is where the Results used by the [MainTable] sheet are entered.

ICP-MS	Al	As	B	Ba	Be
S1	86.38	1.59	94.21	94.21	86.38
S2	87.79	1.66	91.3	88.46	87.79
S3	77.94	1.42	91.88	98.27	77.94
S4	72.11	1.59	95.31	99.41	72.11
S5 Duplicate S3	78.56	1.66	94.69	100.24	78.56
S6	89.84	1.36	85.37	100.91	89.84
S7	89.3	1.69	99.21	81.31	89.3
S8	83.42	1.82	96.33	107.18	83.42
S9	158.97	1.67	92.02	140.1	771

This is an excluded result. Adding "!" turns it into "non-numeric"

Results have a low standard deviation, hence "Low" HorRat.

Outlier?

Typographical error.

Results can be copied from other spreadsheets or databases into this sheet, but care must be taken not to over-write cell formatting - It is recommended that only "clean", unformatted data is pasted. Additional rows can be added by choosing an empty row towards the bottom and adding additional rows beneath it. Excel can also access data from linked external sheets - Users can unlock this table to implement this.

Similarly, it is possible to copy/paste data from the spreadsheet into other files and databases. The default "copy" format is Tab Separated Variable (TSV), which is compatible with most other systems.

---

## MultiAnalytesMU Spreadsheet - [Distributions]

This is the third sheet, and is used by the [MainTable] sheet look up Horwitz Ratio and Grubbs' test values. It should not be edited.

---

## MultiAnalytesMU Spreadsheet - [Units]

This is the fourth sheet, and is used by the [MainTable] sheet. It contains concentration units covering the range from ~10% to parts per-trillion (ppt). If required additional units may be added; but as noted in the sheet, the "Select Units (ppm, µg/kg, g/L..)" selection list in the [MainTable] must be updated to match.

---

## References & Small Print

{1} NATA, Specific Accreditation Criteria ISO/IEC 17025 Application Document Life Sciences - Appendix, Effective: February 2021

and, Neil Shepherd, NATA WA Meeting May 2018, and NATA Webinars November and December 2017

{2} Anal. Chem. 1982, 54, 1, 67–76 Publication Date: January 1, 1982 <https://doi.org/10.1021/ac00238a002>

and, Albert R, Horwitz W (1997) A Heuristic derivation of the horwitz curve. Anal Chem 69:789–790

and, M Thompson AMC Technical Brief No.17 July 2004 - Royal Society of Chemistry 2004

and, M Haustein (on behalf of DAPA), FAO & WHO, Specifications for Pesticides CIPAC Symposium Athens, 2015

{3} Frank E. Grubbs. "Sample Criteria for Testing Outlying Observations." Ann. Math. Statist. 21 (1) 27 - 58, March, 1950. <https://doi.org/10.1214/aoms/1177729885>

*This spreadsheet is "Public Domain". Anyone is free to copy, modify, publish, use, or distribute it for any purpose, commercial or non-commercial, and by any means.*

*The author carries out "Pro Bono" work only; and does not accept any responsibility for any loss caused by the use of, or failure to use, this spreadsheet.*

*Please leave the version number, my name, and copyright information at the bottom of this page.*